

T1 Comparison of Assays Used to Measure Islet Autoantibodies

	SENSITIVE	SPECIFIC	BINDING OCCURS IN FLUID-PHASE	REQUIRES SMALL SAMPLE VOLUME	DOES NOT REQUIRE RADIOACTIVITY	CAN BE MULTIPLEXED
ELISA		+			+	+
Radioimmunoassay	++	++	+			
Bridge-ELISA	+	+	+		+	+
ECL Detection	++	++	+	+	+	+